

Answers to the questions received on 18.02.2013 regarding the tender regulation LUCFI 2013/2/ERDF “Delivery of the Multifunctional Cluster Plant for Deposition of Vacuum Coatings”

Atbildes uz 18.02.2013. saņemtajiem piegādātāja jautājumiem par iepirkuma LU CFI 2013/2/ERAF „Daudzfunkcionālas klāstera iekārtas vakuuma pārklājumu izgatavošanai piegāde” nolikuma prasībām un par piedāvājuma noformēšanu

No. N.p.k.	Question Piegādātāja jautājums	Answer Iepirkumu komisijas atbilde
1	<p>For processing chambers: 2nd and 3rd: The through distance source to substrate is much too short as requested. While we can manufacture the system with this spec, this is not correct technologically. We suggest 250-300mm.</p>	<p>For the 2nd processing chamber the through distance from the evaporator up to the sample can be in the following range: from 100 mm to 250 mm. For the 3rd processing chamber the through distance from the evaporator up to the sample can be in the following range: from 100 mm to 300 mm. Procesu kamerai „2” attālums no tvaicētāja līdz paraugam var būt robežās no 100 mm līdz 250 mm. Procesu kamerai „3” attālums no tvaicētāja līdz paraugam var būt robežās no 100 mm līdz 300 mm.</p>
2	<p>Size and capacity of crucibles shall not be part of the spec as this is a consumable item which can always be replaced. We suggest take out that requirement.</p>	<p>Capacity of crucibles (small: not larger than 1.2 cm<sup>3</sup>) is given in order that the tenderer can offer the crucible holders of the corresponding size. Iztvaicēšanas tīģeļa izmērs (mazs: nedrīkst pārsniegt 1.2 cm<sup>3</sup>) ir dots, lai pretendents varētu piedāvāt atbilstoša izmēra tīģeļa turētājus.</p>